



ESR-300A

Ophthalmic unit



OPHTHALMOLOGY / ENT



Pulse Healthcare Technology
House: 11/1, Shahid Minar Road
Kallyanpur, Dhaka-1216, Bangladesh.
Mobile: +8801708008061
e-mail: info@pht.com.bd
web: www.pht.com.bd



ESR-300A

Ophtalmic unit



Specification

Mechanical Parameters

Left/Right angle of swing arm: $\pm 30^\circ$

Angle of rotation for swing arm right-angle rack: 360°

Angle of rotation for Table: 90°

Slide-board Slippage Range: 500mm

Lifting height for chair: 150~200mm

Loading for chair: 70~150kgs

Electrical Parameters

Input voltage: 110V/220V AC 50HZ~60HZ

Engine power: 55W 24VDC 40W 30VDC

Light luminaire: 18W 250V

Fuse: 5A 250V AC

Power without load: 1W

Pressure test: 2KV

Main wire: 10A 250V

Working mode: S1

Working temperature: $-30^\circ \sim +50^\circ$

Relative humidity: less than 90%

Force of pressing key: 1~5N

Cooling mode: Nature cooling

Installation Dimension

Figure size: 1060mm×1000mm×950mm

Weight: 210kgs



Pulse Healthcare Technology
House: 11/1, Shahid Minar Road
Kallyanpur, Dhaka-1216, Bangladesh.
Mobile: +8801708008061
e-mail: info@pht.com.bd
web: www.pht.com.bd



ESSE3 srl, Via Garibaldi 30
14022 Castelnuovo D.B. (AT)
Tel +39 011 99 27 706
Fax +39 011 99 27 506
e-mail esse3@chierinet.it
web: www.esse3-medical.com

