



MIR-200

Touch Screen Coagulation Analyzer



LABORATORY



Pulse Healthcare Technology

House: 11/1, Shahid Minar Road
Kallyanpur, Dhaka-1216, Bangladesh.

Mobile: +8801708008061

e-mail: info@pht.com.bd

web: www.pht.com.bd

MIR 200

Touch Screen Coagulation Analyzer



- *Double magnetic route*
- *Auto-linked pipette*
- *Testing channel: 2*
- *7" color Touch Screen*
- *Built-in thermal printer*

Main Features

- Double magnetic route, stable performance without interference
- The CPU adopts ARM technology to measure the motion of the magnetic ball to work out the coagulation time
- It can perform multi-parameter simultaneously with better accuracy, precision and repetition
- 100 assay parameters and 1000 test result can be saved
- Clock CMOS chip inserted
- Reagent open system
- Built-in thermal printer 58mm
- Interface: RS232, USB, pipette-linked port
- Power supply: AC220V $\pm 15\%$, 50Hz, 80VA

Technical features

Testing channels	2
Test method	double magnetic route
Main tested parameters	PT, APTT, TT, FIB
Other tested parameters	Hep, ProC, ProS, Coagulation factor, II, V, VII, VIII, IX, X, XI, XII, AT-III, AT, RepT, RVVT, RVVCT, RVVRT, LA
Sample consumption	$\leq 200\mu\text{l}$
Reagent consumption	$50\mu\text{l} \sim 200\mu\text{l}$
Temperature	$37^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$
Adding sample	by a timing linkage
Precision	$\text{CV} \leq 3\%$
Test speed	$\leq 300\text{s}$ (depend on setting)
Sample warming-up	18/8 positions
Reagent warming-up	6/6 positions
Dimension (LxWxH)	370 x 290 x 120 mm
Weight	4 kg (net) ; 6 kg (gross)



Pulse Healthcare Technology
 House: 11/1, Shahid Minar Road
 Kallyanpur, Dhaka-1216, Bangladesh.
 Mobile: +8801708008061
 e-mail: info@pht.com.bd
 web: www.pht.com.bd



ESSE3 srl, Via Garibaldi 30
 14022 Castelnuovo D.B. (AT)
 Tel +39 011 99 27 706
 Fax +39 011 99 27 506
 e-mail esse3@chierinet.it
 web: www.esse3-medical.com

